



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

SOUTHWEST RESEARCH INSTITUTE  
Office of Automotive Engineering  
Fuels and Lubricants Research Division and the  
Powertrain Engineering Division  
6220 Culebra Street, P.O. Box 28510  
San Antonio, TX 78228-0510  
Paul Nemeth Phone: 210 522 5891

CALIBRATION

Valid To: May 31, 2026

Certificate Number: 0702.05

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Particle Size & Number Measurement Systems Calibration

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Particle Diameter	60 nm, 100 nm	1.1 nm	ISO 15900 (SwRI TIP 06-080)
	Cardinal Points	1.2 nm	
Particle Detection Efficiency	30 % to 110 % of efficiency	7.5 % of efficiency	ISO 27891 (SwRI TIP 06-079)
Particle Reduction Factor	(1 to 10 000) (unitless)	11 (unitless)	UN-ECE R49 & R83 (SwRI TIP 06-082)
Particle Removal Efficiency	0 % to 100 % of removal efficiency	0.16 % of removal efficiency	UN-ECE R49 & R83 (SwRI TIP 06-082)
		0.11 % of removal efficiency	SAE AIR6241 (SwRI TIP 06-081)

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Particle Penetration	0 % to 100 % penetration of particles	11 % penetration of particles	SAE AIR6241 (SwRI TIP 06-081)

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.





# Accredited Laboratory

A2LA has accredited

## SOUTHWEST RESEARCH INSTITUTE

*San Antonio, TX*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 3<sup>rd</sup> day of July 2024.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 0702.05  
Valid to May 31, 2026

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*